

# HITACHI

 Hitachi Medical Corporation

MEDICAL SYSTEM OPERATIONS GROUP, KASHIWA  
2-1 SHINTOYOFUTA, KASHIWA-SHI, CHIBA 277-0804, JAPAN

## EC-CONFORMITY DECLARATION

### Manufacturer

Hitachi Medical Corporation  
4-14-1, Soto-Kanda, Chiyoda-ku  
Tokyo, 101-0021, Japan

### European Representative

Hitachi Medical Systems GmbH  
Kreuzberger Ring 66  
D-65205 Wiesbaden, Germany

### Medical Device

Modality : Transducer, Ultrasonic  
UMDNS Code : 14-121  
MDD Classification : IIa (Rule 10)  
Model Name : Convex Array Probe  
Model Type : EUP-CV724  
The first manufactured serial number: PE12627


This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaration and is only valid in connection with a batch specific Certificate of Compliance for all products concerned bearing the CE mark, as documented in each in-process check list of the form No. SKM C2-191 for manufacturing.

The undersigned hereby declares that the medical device as specified above and related options comply with the essential requirements of Annex I of the EC-Directive 93/42/EEC.

The declaration of conformity is based on an assessment procedure in compliance with the EC Directive 93/42/ EEC, Annex II for a Full Quality Assurance System.

Notified Body: TÜV Rheinland LGA Products GmbH  
Tillystraße 2, 90431 Nürnberg  
Germany

Ident No.: CE-0197



K. Kuriyama  
General Manager  
Ultrasound System Division



N. Hisamoto  
Division Manager  
Quality Assurance Division

Place JAPAN

Date 2011-8-28